



**Check that the Coolsense cartridge contains alcohol before using the device! Ensure that the metal pin is coated with the alcohol gel prior to use!**

**Failure to do so may cause a cold burn.**

**\*Replace with a new cartridge if empty**



Distributed by Balance Medical

[www.balancemedical.com.au](http://www.balancemedical.com.au), phone: 1300 668 099